



INNOVATIVE IDEAS
EXCEPTIONAL DESIGN
UNMATCHED CLIENT SERVICE

TRANSMITTAL LETTER

DATE: January 24, 2024

TO: City of Detroit
Environmental Affairs
Buildings, Safety Engineering and Environmental Department
2 Woodward Avenue
Detroit, Michigan 48226
Phone: 313.471-5110
Attn: Mr. Hossam N. Hassanien, PG, CPG
Email: hassanienh@detroitmi.gov

RE: Former State Fair Grounds Ambient Monitoring Second Construction Phase Report

PROJECT # 2142-7261-00

WE ARE TRANSMITTING HERewith THE FOLLOWING MATERIAL

Date	Copies	Description
01/24/2024	1	Former State Fair Grounds Ambient Monitoring Second Construction Phase Report

REMARKS

Please find attached a copy of the Ralph C. Wilson Geotechnical Monitoring Report for your use. If you have any questions, you may contact Mr. Dor'Mario Brown at 248.727.7083. Thank you.

DLZ REPRESENTATIVE

Dor'Mario Brown
Division Manager



INNOVATIVE IDEAS
EXCEPTIONAL DESIGN
UNMATCHED CLIENT SERVICE

January 24, 2024

Ms. Donna Rice
City of Detroit
Detroit Building Authority
500 Griswold, Suite 200
Detroit, Michigan 48226

RE: Ambient Air Quality Monitoring – 2nd Construction Phase Ambient Monitoring Report
Proposed Department of Transportation (DDOT) Transit Center
Detroit, Michigan
Project No. 2142726100

Dear Ms. Rice:

The City of Detroit Department of Transportation (DDOT) recently completed a property transaction for a new Transit Center to be constructed on Parcel D of the former Michigan State Fairgrounds located at 8 Mile Road and Woodward Avenue in Detroit, Michigan. The City contracted DLZ Michigan, Inc. to conduct ambient air quality monitoring at the proposed Detroit Department of Transportation (DDOT) Transit Center site (Site).

The monitoring program consists of siting localized monitors at an upwind and downwind locations to measure concentrations of particulate matter (PM₁₀ and PM_{2.5}), nitrogen oxide (NO_x, as NO₂), and volatile organic compounds (VOCs), and evaluate air quality from the Site during three (3) distinct phases:

- Pre-development baseline period
- Construction phase
- Post-construction facility operation

Pre-Development Baseline Period (Completed)
1st Construction Phase Monitoring (Completed)
2nd Construction Phase Monitoring (Completed)

DLZ's 2nd Construction Phase Monitoring report, dated November 20, 2023, presented ambient concentrations during the construction activities at the Site. 2nd Construction phase period included monitoring data collected by Montrose Air Quality Services, LLC (MAQS), from September 15 through September 23, 2023, and was supplemented with monitoring data collected by the Michigan Department of Environment, Great Lakes, and Energy (EGLE) from September 15, 2023, through September 23, 2023. The purpose of the Construction Monitoring Report was to compare the data collected at the Site during the construction activities to corresponding NAAQS and baseline reference concentrations. If measured pollutant concentrations exceeded, the construction contractor would be alerted to investigate on-site construction

607 Shelby St., Ste. 650, Detroit, Michigan 48226 | OFFICE 313.961.4040 | ONLINE WWW.DLZ.COM

Akron Bellefontaine Bridgeville Burns Harbor Chicago Cincinnati Cleveland Columbus Detroit Fort Wayne Grand Rapids Indianapolis Joliet
Kalamazoo Knoxville Lansing Lexington Logan Madison Maumee Melvindale Merrillville Munster Muskegon Port Huron Saint Joseph San
José South Bend Waterford

activities at the time that the elevated concentration was recorded and determine if additional mitigation measures were needed to reduce pollutant concentrations to below the baseline reference concentration.

2nd Phase Construction Monitoring

The enclosed report presents the results of the 2nd Construction Phase Monitoring event that was conducted for the two (2)-weeks period of September 15, 2023, through September 23, 2023. The goal of Construction Phase Ambient monitoring is to collect concentration data of target air pollutants during the on-site activities consisting of concrete work, steel construction, roofing, interior buildout, electrical work, and plumbing to assess whether additional mitigation efforts are warranted to reduce pollutant concentrations to below baseline levels or NAAQS.

The enclosed 1st Phase Construction Ambient Monitoring Report describes the monitoring program, objectives, Site overview, monitor locations and equipment, monitoring results, and an overview of data quality assurance.

The report includes monitoring data from two (2) available sources, including:

- Two (2) Site monitors operated by MAQS for DLZ during the monitoring period (September 15, 2023 through September 23, 2023) and identified as Unit 1838 (upwind location) and Unit 1839 (downwind location).
- Nearby off-site monitors operated by Michigan Department of Environment, Great Lakes, and Energy (EGLE) during the MAQS monitoring period.

As part of this air monitoring program, MAQS collected two (2) weeks of air monitoring data for NO_x (as NO₂), PM₁₀ and PM_{2.5}, and VOCs at two (2) monitors, along with prevailing wind directions and speeds (vectors).

The City anticipates that development of the proposed DDOT Transit Center may result in direct and fugitive air emissions from construction activities, as well as future operations. Sources of NO_x and VOC emissions related to construction may include vehicular traffic and diesel engines (over-the-road and non-road heavy duty construction). Potential emissions of PM₁₀ and PM_{2.5} related to construction may include fugitive dust associated with vehicular traffic, soil handling, material storage piles, concrete batching, and abrasives blasting.

The monitors, designated as Unit 1838 and Unit 1839, were located on opposite sides of the Site and both stations are configured to collect pollutant and meteorological data. The upwind monitor (Unit 1838) measures pollutant concentrations that have not blown across the Site and should be free from potential impacts of on-site development activity and is representative of local area background concentrations.

Results of 2nd Phase Construction Monitoring

As presented below and in the enclosed report, for monitoring conducted September 15 through September 23, 2023, concentrations of PM₁₀, PM_{2.5}, NO_x (as NO₂) and VOC from the on-site monitors are establishing their baseline concentrations, as summarized in Table 2. NO_x (as NO₂) concentrations are less than the 1-hour NAAQS of 100 ppb for NO₂.¹ Monitored concentrations of PM₁₀, PM_{2.5} are also less than the 24-hour NAAQS of 150 µg/m³ for PM₁₀, 35 µg/m³ for PM_{2.5}.

Table 2 – Summary of Air Monitoring from September 15 through September 23, 2023

Pollutant	2nd phase Maximum Concentration	2nd Phase Max Monitor	Date of Maximum Concentration	Baseline Max Concentration	Baseline Max Monitor	NAAQS	Units
PM ₁₀	35.7	Unit 1839	9-20-2023	17	Unit 1839	150	µg/m ³
PM _{2.5}	9.2	Unit 1839	9-20-2023	4	Unit 1839	35	µg/m ³
NO ₂	36	Unit 1838	9-22-2023	22	Unit 1838	100	ppb
VOC	0.06	Unit 1838	9-16-2023	0.03	Unit 1839	NA ²	ppm

¹ Construction Phase Monitoring report included two (2) Site monitors operated by MAQS for DLZ from July 15, 2023, through July 24, 2023, and identified as Unit 1838 (upwind location) and Unit 1839 (downwind location), as well as monitoring data provided by Michigan Department of Environment, Great Lakes, and Energy (EGLE).

² NAAQS have not been established for VOC. VOCs are considered precursors to the formation of ozone. Ozone is formed downwind by photochemical reaction of NO_x and VOCs in certain ambient conditions (typically hot, sunny weather)

In summary, the data collected from the site during the first construction-phase do not exceed any NAAQS. However, during this monitoring period, there were periods where ambient concentrations of PM₁₀, PM_{2.5}, NO₂ and VOC shows elevated as compared to the baseline concentration, this is possibly due to the construction activities at the DDOT transit center site. To conclude the data collected are not indicative of a threat to public health.

We appreciate this opportunity to be of service to you. If you have questions or need additional information, please contact us at 248-727-7083.

Sincerely,
DLZ Michigan, Inc.



Dor'Mario Brown
Division Manager
DB/GS

2ND CONSTRUCTION PHASE AMBIENT MONITORING REPORT DDOT TRANSIT CENTER AT FORMER MICHIGAN STATE FAIRGROUNDS DETROIT, MICHIGAN

Prepared For:

DLZ Michigan, Inc.

607 Shelby St. Suite 650

Detroit, MI 48226

Prepared By:

Montrose Air Quality Services, LLC

45 U.S. 46, Suite 601

Pine Brook, NJ 07058

4949 Fernlee Avenue

Royal Oak, MI 48073

Document Number: **027AA-016697-RT-84**

Monitoring Period: **September 15 through September 23, 2023**

Submittal Date: **November , 2023**

Table of Contents

PROJECT OVERVIEW	2
<i>Background</i>	<i>2</i>
<i>Objectives</i>	<i>2</i>
<i>Potential Sources.....</i>	<i>2</i>
<i>Site Overview.....</i>	<i>3</i>
<i>Figure 1 – Monitor Locations at the DDOT Transit Center Construction Site</i>	<i>4</i>
<i>Monitoring Equipment</i>	<i>5</i>
<i>Discussion of Results.....</i>	<i>6</i>
POLLUTANT DATA COLLECTED	7
<i>Figure 2 – PM₁₀ Data.....</i>	<i>7</i>
<i>Figure 3 – PM_{2.5} Data</i>	<i>8</i>
<i>Figure 4 – NO₂ Data</i>	<i>9</i>
<i>Figure 5 – VOC Data.....</i>	<i>10</i>
METEOROLOGICAL DATA COLLECTED	11
<i>Figure 6 – Wind Rose From 1838 Monitor.....</i>	<i>11</i>
<i>Figure 7 – Wind Rose From 1839 Monitor.....</i>	<i>12</i>
DATA QUALITY ASSURANCE/QUALITY CONTROL	13
<i>Quality Assurance/Quality Control.....</i>	<i>13</i>
CONCLUSION	14
SIGNATURE PAGE.....	15
APPENDICES	16
Appendix A: Quality Assurance Logs.....	16
Appendix B: Calibration Certification Sheets	17
Appendix C: Locations of MI EGLE Monitors Relative to the Former State Fairgrounds.....	18

Project Overview

Background

DLZ Michigan, Inc. (DLZ) has retained Montrose Air Quality Services, LLC (Montrose) to conduct an ambient air monitoring program in support of the proposed Detroit Department of Transportation (DDOT) Transit Center on Parcel D of the former Michigan State Fairgrounds located at 8 Mile Road and Woodward Avenue in Detroit, Michigan. The program is conducted to monitor for a mixture of pollutants that may originate from construction activities as well as future Site operations including vehicular traffic, surface attrition, and dust emissions.

The previously-submitted Baseline Monitoring Report presented ambient monitoring data collected by Montrose prior to commencement of significant Site construction activities. The baseline monitoring period began July 8 and continued through July 22, 2022. The purpose of the Baseline Monitoring report is to characterize background ambient concentrations at the Site for each monitored pollutant. The pollutant concentrations recorded during the Baseline monitoring period were quite low. Consequently, Montrose selected the highest hourly concentration recorded for NO₂ and VOC during the baseline monitoring period to determine reference baseline values for NO₂ and VOC. Similarly, Montrose selected the highest 24-hour averaged concentration recorded for PM_{2.5} and PM₁₀ during the baseline monitoring period to determine reference baseline values for PM_{2.5} and PM₁₀. It should be noted that the resulting baseline reference concentrations are far below National Ambient Air Quality Standards (NAAQS) established for NO₂, PM_{2.5} and PM₁₀.

Data collected at the Site during subsequent construction and post-construction monitoring periods are compared to corresponding NAAQS and baseline reference concentrations. For construction-phase monitoring periods, if measured pollutant concentrations exceeded the NAAQS concentration and corresponding meteorological (i.e., wind) data indicated the elevated concentration might have resulted from on-site activity (as opposed to transport from off-site sources), the construction contractor would be alerted to investigate on-site construction activities at the time that the elevated concentration was recorded and determine if additional mitigation measures were needed to reduce pollutant concentrations to below the baseline reference concentration.

This report also includes data reported from air pollutant monitors operated by Montrose and Michigan Department of Environment, Great Lakes, and Energy (EGLE) during the monitoring period commencing on July 15 and concluding on July 24, 2023.

Objectives

The specific objectives are to continuously measure ambient concentrations of the following pollutant and meteorological parameters at two (2) locations proximate to the Site:

- Suspended particulate matter having an aerodynamic diameter ≤ 10 microns (PM₁₀)
- Suspended particulate matter having an aerodynamic diameter ≤ 2.5 microns (PM_{2.5})
- Nitrogen Dioxide (NO₂)
- Volatile Organic Compounds (VOC)
- Meteorological parameters measured at each monitoring location: wind speed, wind direction, temperature, relative humidity, and barometric pressure

Potential Sources

Sources of NO_x and VOC emissions related to construction include vehicular traffic and diesel engines (over-the-road and non-road heavy duty construction). Potential emissions of PM₁₀ and PM_{2.5} related to construction may include fugitive dust associated with vehicular traffic, soil handling, material storage piles, concrete batching, and abrasives blasting.

Operational Staff and Contacts

Facility Information

Monitoring Proposed DDOT Transit Center
Location: former Michigan State Fairgrounds
1120 W. State Fair Avenue
Detroit, MI 48203

Monitoring Program Coordinator

DLZ Michigan, Inc.
607 Shelby St., Suite 650
Detroit, MI 48226

Project Contacts: Mr. Dor'Mario Brown
Role: Division Manager
Company: DLZ Michigan, Inc.
Telephone: 313-383-3216
Email: dbrown@dlz.com

Monitoring Team Contact Information

Testing Firm: Montrose Air Quality Services, LLC (Montrose)

Contact: David Cummings	Darrin Barton
Title: District Manager	Sr. Project Manager
Telephone: 201-213-2913	512-656-6455
Email: dcummings@montrose-env.com	dabarton@montrose-env.com
Kevin Ruggiero	Jeffrey Peitzsch
Sr. Project Manager	Shop Coordinator
973-417-6487	313-213-4816
kruggiero@montrose-env.com	jbpeitzsch@montrose-env.com
Linda Quigley	
Senior Reporting QC Specialist	
973-202-3312	
lquigley@montrose-env.com	

Site Overview

The air quality monitoring is performed at the site of the proposed DDOT Transit Center (former Michigan State Fairgrounds) property located at 1120 W State Fair Avenue in Detroit, MI. The existing site contains historically significant buildings: the 1924 Coliseum, the 1926 Dairy Cattle Building and the adjacent Agricultural Building. These structures may be retained or reused for the Transit Center. Other structures onsite in this area will be demolished and re-used to build a new DDOT Transit Center. Figure 1 presents an aerial view of the Site showing the DDOT Transit Center construction site and locations of the upwind (#1838) and downwind (#1839) air quality monitors.

Figure 1 – Monitor Locations at the DDOT Transit Center Construction Site



Monitoring Equipment

Air monitoring at the proposed DDOT Transit Center (former Michigan State Fairgrounds) is performed using an AQS-1 Urban Air Quality Monitor manufactured by Aeroqual. The compact size of the AQS-1 monitor makes it viable for a changing construction site where the monitor equipment may need to be removed and re-deployed during monitoring campaigns. Air monitoring is conducted for the parameters listed in Table 1.

Table 1 - Pollutants Monitored

Air Pollutant/Parameter Category	Principle of Operation
PM ₁₀ and PM _{2.5}	Laser Scattering interferometry with particle counting
NO ₂	Electrochemical
VOC	Photoionization
Wind Speed, Wind Direction, Temperature, Relative Humidity, Barometric Pressure	Sonic Anemometer and Various

The AQS-1 integrates all measurement detectors, sample pump, flow controllers, signal processing, data acquisition and data transmission components within a compact, weatherproof enclosure. The AQS-1 features separate, dedicated sample air inlets configured specifically for the measurement of suspended particulate matter (i.e., PM₁₀ and PM_{2.5}) and gaseous pollutants (i.e., NO₂ and VOC). An internal sample pump and flow controllers regulate and maintain stable, optimal flow rates of ambient air through each sample inlet. The sample air streams are directed to the various detection and measurement modules housed within the instrument. Each AQS monitor is powered in the field by deep-cycle batteries charged via solar photovoltaic panels and a battery charging regulator.

Particulate matter is continuously measured via laser scattering interferometry and particle counting methodology. This method is based on the physical principle of light scattering. Each single particle in the detection and measurement module is illuminated by a defined laser light beam; the coherent laser light is scattered by reflection off particles in the sample air stream within the detector. The scattering signal is detected at an angle of 90° by a photo diode within the detector module. In accordance with the Mie theory, each measured pulse height of the scattered light is directly proportional to the particle size. The pulses are classified in an electronic register of 32 different size channels.

NO₂ is continuously measured using an electrochemical sensor consisting of a working counter and reference electrode. NO₂ concentrations are detected and measured by oxidation or reduction reactions on an electrochemical sensor housed within a module containing a liquid electrolyte specific to NO₂. The electrochemical sensor is subjected to a controlled, external electrical circuit. When NO₂ is present, a current proportional to the NO₂ concentration is produced.

VOC is continuously measured using a photoionization detector (PID). The PID sensor lamp produces photons having enough energy to ionize VOC molecules. The PID will only respond to molecules that have an ionization energy at or below the energy of the lamp; the PID used in the AQS-1 project employs a 10.6 electron-volt lamp. The ions produced from VOC compounds generate an electrical current that is measured as the output of the detector.

The meteorological monitors integrated with the AQS-1 are the Vaisala Model WXT536 Weather Transmitter. The meteorological monitors are mounted on a rigid support post elevated above the monitor enclosure cabinet, and are integrated with the data acquisition and data telemetry system housed within the PM_{2.5} monitor enclosure.

Measurement signals produced by each pollutant detector and the meteorological monitors are acquired by an internal mini-computer that processes, scales, averages and stores the measurement data. The internal computer is integrated with a wireless (cellular service) data modem that supports bidirectional communications.

Monitoring methods and activities employed in the monitoring program, including instrument calibration, operation, maintenance and quality control (QC) activities, were performed in accordance with the protocols and procedures contained in the approved *Ambient Air Test Plan 2022 Proposed DDOT Transit Center at Former Michigan State Fairgrounds* dated June 17, 2022.

Discussion of Results

The results of PM₁₀, PM_{2.5}, NO₂, and VOC monitoring data are presented in Figures 2 through 5 in this report. These figures also include data reported from nearby air monitoring stations maintained by the Michigan Department of Environment, Great Lakes, and Energy (EGLE) for the same time period. The EGLE data contained in this report are from monitors that are routinely subjected to calibration and maintenance. It should be noted that, as of the date of this report, the EGLE data have not yet been processed through EGLE final quality assurance procedures. The monitor locations for EGLE Sites can be found on the map provided in Appendix C (*Locations of MI EGLE Monitors Relative to the Former State Fairgrounds*).

The Clean Air Act requires EPA to establish National Ambient Air Quality Standards (NAAQS) for certain air pollutants considered harmful to public health and the environment. Air pollutants for which NAAQS are established include NO₂, PM_{2.5} and PM₁₀. NAAQS have not been established for VOCs. VOCs are considered precursors to the formation of ozone. Ozone is formed by photochemical reactions of NO_x and VOCs in certain ambient conditions.

The graphed data shown in Figures 2 through 5 present measured concentrations for these pollutants collected during the monitoring period relative to the Baseline concentration and corresponding NAAQS.

The NAAQS for NO₂, PM_{2.5}, and PM₁₀ were not exceeded during these monitoring periods.

Electronic records of all data and calibrations have been uploaded to the Montrose Data Server, where they will be archived for a period of at least three (3) years.

Pollutant Data Collected

Figure 2 –PM₁₀ Data

Figure 2 below presents the ambient PM₁₀ measurement data collected at the DDOT Transit Center construction site on Parcel D of the former Michigan State Fairgrounds property during the monitoring period of 9/15/23 to 9/23/23. This graph is a plot of the PM₁₀ measurement data as averaged over each 24-hour day (midnight-to-midnight) during the monitoring period. The PM₁₀ daily averaging interval used for this monitoring program is consistent with the EPA 24-hour averaging interval used for NAAQS data reporting assessments. The primary and secondary PM₁₀ NAAQS is equal to a daily averaged value of 150 micrograms per cubic meter (µg/m³) not to be exceeded more than once per year on average over 3 years.

The solid yellow line represents in Figure 2 below represents the 24-hour PM₁₀ NAAQS of 150 µg/m³. The solid red line represents the baseline PM₁₀ concentration of 15.7 µg/m³ derived from the Baseline monitoring interval. The additional graphed data in Figure 2 presents 24-hour averaged PM₁₀ data reported from each of the on-site monitors as well as corresponding data reported from the MI EGLE Dearborn continuous PM₁₀ monitor, which is the closest state-operated PM₁₀ monitor relative to the former Michigan State Fairgrounds property.

Figure 2: PM₁₀ Data




Figure 3 –PM_{2.5} Data

Figure 3 below presents the ambient PM_{2.5} measurement data collected at the DDOT Transit Center construction site on Parcel D of the former Michigan State Fairgrounds property during the monitoring period starting on 9/15/23 and ending on 9/23/23. This graph is a plot of the PM_{2.5} measurement data as averaged over each 24-hour day (midnight-to-midnight) during the monitoring period. The PM_{2.5} daily averaging interval used for this monitoring program is consistent with the EPA 24-hour averaging interval used for NAAQS data reporting assessments. The primary and secondary PM_{2.5} NAAQS is equal to a daily averaged value of 35 micrograms per cubic meter (µg/m³) not to be exceeded more than once per year on average over 3 years.

The solid yellow line Figure 3 below represents the 24-hour PM_{2.5} NAAQS of 35 µg/m³. The solid red line represents the baseline concentration of 3.8 µg/m³ derived from the Baseline monitoring interval. The additional graphed data in Figure 3 presents 24-hour averaged PM_{2.5} data reported from each of the on-site monitors as well as corresponding data reported from the MI EGLE Dearborn and Detroit SW PM_{2.5} monitors, which are the closest state-operated continuous PM_{2.5} monitors relative to the former Michigan State Fairgrounds property. (Note: The MI EGLE also operates a PM_{2.5} monitor at the Oak Park monitoring site, which is located closer to the former Michigan State Fairgrounds property. The Oak Park PM_{2.5} monitor collects filter-based PM_{2.5} samples at 3-day intervals. Laboratory analytical results for filter-based PM samples are not available until approximately three months after the sample date. Consequently, the MI EGLE Oak Park PM_{2.5} data are not available for inclusion in this report.)

Figure 3: PM_{2.5} Data




Figure 4 – NO₂ Data

Figure 4 below presents the ambient NO₂ measurement data collected at the DDOT Transit Center construction site on Parcel D of the former Michigan State Fairgrounds property during the monitoring period starting on 7/15/23 and ending on 7/24/23. This graph is a plot of the NO₂ measurement data as averaged over one (1) hour intervals. This is consistent with the associated EPA primary NO₂ NAAQS: A 1-hour averaged value of 100 parts-per-billion (ppb) not to be exceeded more than once per year on average over 3 years.

The solid yellow line in Figure 4 represents the 1-hour NO₂ NAAQS of 100 ppb. The solid red line represents the baseline NO₂ concentration of 25.6 ppb derived from the Baseline monitoring interval. The additional graphed data in Figure 4 presents the 1-hour averaged data NO₂ data reported from each of the on-site monitors as well as corresponding data reported from the MI EGLE Detroit SW continuous NO₂ monitor, which is the closest state-operated NO₂ monitor relative to the former Michigan State Fairgrounds property.

Figure 4: NO₂ Data





Figure 5 –VOC Data

Figure 5 below presents the ambient VOC measurement data collected at the DDOT Transit Center construction site on Parcel D of the former Michigan State Fairgrounds property during the monitoring period starting on 7/15/23 and ending on 7/24/23. The US EPA does not promulgate a NAAQS for VOC.

The solid red line in Figure 5 represents the baseline hourly-averaged VOC concentration of 0.03 parts-per-million (ppm) derived from the Baseline monitoring interval. The additional graphed data in Figure 5 presents the 1-hour averaged data VOC data reported reported from each of the on-site monitors. MI EGLE does not monitor for VOC at any nearby MI EGLE monitoring sites. Consequently, no meaningful MI EGLE VOC data are available for comparison purposes.

Figure 5: VOC Data



Meteorological Data Collected

Figure 6 presents a wind rose derived from the wind speed and wind direction data collected from AQS-1 Upwind Monitor (S/N 1838) over the course of the monitoring period of 9/15/23 to 9/23/23. AQS-1 Monitor was deployed at a nominally upwind location at the DDOT Transit Center construction site, as depicted in Figure 1 in this report.

Figure 6: Wind Rose From AQS-1 (1838) Upwind Meteorological Monitor





Figure 7 presents a wind rose derived from the wind speed and wind direction data collected from AQS-2 Downwind Monitor (S/N 1839) over the course of the monitoring period of 9/15/23 to 9/23/23. AQS-2 was deployed at a nominally downwind location at the DDOT Transit Center construction site, as depicted in Figure 1 in this report.

Figure 7 – Wind Rose From AQS-2 (1839) Downwind Meteorological Monitor



As is evident from the wind rose data, winds from the south, southwest, and west were predominant during the monitoring period of 9/15/23 to 9/23/23. Wind speeds recorded were also predominantly light, being mostly within the range of 0.5 to 3.6 m/s.

Data Quality Assurance/Quality Control

Quality Assurance/Quality Control

Quality assurance is a general term for the procedures used to ensure that a particular measurement meets the quality requirements for its intended use. Quality control for monitoring instrumentation consists of calibrations, sample flow rate verifications, leak checks and verification of other monitor performance indicators.

Monitoring methods and activities employed in the monitoring program, including instrument calibration, operation, maintenance and quality control (QC) activities, were performed in accordance with the protocols and procedures contained in the approved *Ambient Air Test Plan 2022 Proposed DDOT Transit Center at Former Michigan State Fairgrounds* dated June 17, 2022.

All quality control data for the on-site monitors operated at the former Michigan State Fairgrounds property can be found in Appendix A to this report, entitled “*Quality Assurance Logs*”. Certificates of traceability for the calibration standards and equipment used in support of quality assurance checks are presented in Appendix B to this report entitled “*Calibration Certification Sheets*”.

Conclusion

The ambient air quality monitoring data collected from the site during the second DDOT Transit Center construction-phase monitoring period of September 15 to September 23, 2023 do not exceed any NAAQS. During this monitoring period, the on-site monitors and nearby MI EGLE monitors all recorded periods during which ambient concentrations of PM_{2.5}, PM₁₀, NO₂ and VOC were elevated beyond the baseline concentrations. These elevated concentrations may be attributed in part to construction activities at the DDOT Transit Center site.

Signature Page

This report was prepared and reviewed by the following individuals:



Linda Quigley
Data Manager
Montrose Air Quality Services, LLC



David Cummings
District Manager
Montrose Air Quality Services, LLC

Appendices

Appendix A: Quality Assurance Logs

AEROQUAL AQS-1 VOC HIGH RANGE MODULE VERIFICATION/CALIBRATION FORM

Network:	City of Detroit (DLZ)	Site:	MTMS Lab	Date:	9/11/23
Time Off-Line:	11:30 EST	Time On-Line:	14:07 EST	Technician:	Jeremy Levine

Calibration Equipment Info.	Analyzer Model:	Aeroqual AQS-1	S/N:	1838	Last Cal:	8/30/23
	Calibrator Model No:	Teledyne API	S/N:	69	Cal. Date:	3/2/23
	Zero Air Model No:	Teledyne API	S/N:	n/a	Cert Date:	n/a
	Gas Supplier:	AirGas	Cyl. Conc. (PPM):	49.33	Cyl. Pressure (PSIG)	2,080

VOC Sensor Module Calibration Settings	"As Found" (Before Any Adjustment)	"As Left" (After Adjustment)
OFFSET	0.00	
GAIN	1.389	

"AS FOUND" (UNADJUSTED) TEST DATA

Calibrator Flow and Test Gas Data					Observed VOC Response from AQS-1		Error (Δ%)
Calibrator Gas Channel		Calibrator Air Channel		Known VOC Input Gas Conc. (PPM)	Response (PPM)	Std. Dev. (PPM)	
Display Setting (SLPM)	Actual Flow Rate (SLPM)	Display Setting (SLPM)	Actual Flow Rate (SLPM)				
OFF	OFF	5.0000	5.0153	0.00	0.00	0.00	-
0.0500	0.0502	4.9493	4.9760	0.49	0.52	0.00	6.1%
0.0500	0.0501	2.4493	2.4680	0.98	0.97	0.00	-1.0%

"AS LEFT" (ADJUSTED) TEST DATA

Calibrator Flow and Test Gas Data					Observed VOC Response from AQS-1		Error (Δ%)
Calibrator Gas Channel		Calibrator Air Channel		Known VOC Input Gas Conc. (PPM)	Response (PPM)	Std. Dev. (PPM)	
Display Setting (SLPM)	Actual Flow Rate (SLPM)	Display Setting (SLPM)	Actual Flow Rate (SLPM)				
OFF	OFF	5.0000		0.00			-

NOTES:

- The VOC sensor zero response should be 0.0 ppm ± 0.2 ppm with a Std. Dev. < 0.2 ppm. If the sensor response error is greater than ± 0.2 ppm then an offset adjustment is required. If the Std. Dev. is greater than 0.2 ppm then the sensor is outside acceptable range and may need replacement.
- The adjusted zero response NEW offset should be -1 < OFFSET < 1 and the sensor response 0.0 ppm ± 0.2 ppm.
- The VOC sensor SPAN response should be ± 1 ppm (5% span of 20 ppm) with a Std. Dev. < 0.4 ppm (2% span of 20 ppm). If the sensor response error is greater than ± 1 ppm then a GAIN adjustment is required. If the Std. Dev. is greater than 0.4 ppm then the sensor is outside acceptable range and may need replacement.
- The adjusted span response NEW gain should be 0.2 < GAIN < 5.0 and the sensor response 0.0 ppm ± 1 ppm.

Comments:

Values good, no adjustment needed.

Technician: Jeremy Levine

QA Review: 

MONTROSE AIR QUALITY SERVICES LLC

AEROQUAL AQS-1 NO2 MODULE MULTI-POINT CALIBRATION FORM

Calibration Data on This Form Are For:				Unadjusted Cal.	X	Adjusted Cal.	
Network:	City of Detroit (DLZ)	Site:	MTMS Lab	Date:	9/11/23		
Time Off-Line:	14:10 EST	Time On-Line:	17:45 EST	Technician:	Jeremy Levine		

Calibration Equipment Info.	Analyzer Model:	Aeroqual AQS-1	S/N:	1838	Last Cal:	8/31/23
	Calibrator Model No.:	Teledyne API	S/N:	69	Cal. Date:	3/2/23
	Zero Air Model No.:	Teledyne API	S/N:	n/a	Cert Date:	n/a
	Gas Supplier:	Airgas	Cyl. Cert. Date:	1/26/21	Cyl. Pressure (PSIG)	1,320
	Gas Cylinder ID #:	D068357	Cyl. Conc. (PPM):	30.95	Gas Module Total Flow Rate	130 mL

Analyzer Calibration Settings	"As Found" (Before Any Adjustment)	"As Left" (After Adjustment)
OFFSET	-0.6	
GAIN	1.213	

Calibrator Flow and Test Gas Data					NO ₂ Response		Δ% (Observed Response Vs. Known Conc.) 3	PASS/FAIL
Calibrator Gas Channel		Calibrator Air Channel		Known NO ₂ Gas Conc. (PPB)	Observed from AQS-1			
Display Setting (SLPM)	Actual Flow Rate (SLPM)	Display Setting (SLPM)	Actual Flow Rate (SLPM)			Response (PPB)	Std. Dev. (PPB)	
0.0452	0.0453	3.4548	3.4788	397.8	390.6	0.2	-1.8%	
0.0484	0.0485	4.9516	4.9793	298.6	295.6	0.5	-1.0%	
0.0323	0.0324	4.9677	4.9933	199.5	197.0	0.5	-1.3%	
0.0161	0.0163	4.9839	5.0110	100.3	97.5	0.3	-2.8%	
OFF	OFF	5.0000	5.0171	0.0	-0.1	0.4	-	

Linear Regression Analysis:					
Slope:	0.985510	Intercept:	-0.232968	Corr. Coefficient (r):	0.999971

NOTES:

- The NO2 sensor zero response should be 0.0 ppb ± 0.2 ppb with a Std. Dev. < 0.2 ppb. If the sensor response error is greater than ± 0.2 ppb then an offset adjustment is required. If the Std. Dev. is greater than 0.2 ppb then the sensor is outside acceptable range and may need replacement.
- The adjusted zero response NEW offset should be -1 < OFFSET < 1 and the sensor response 0.0 ppb ± 0.2 ppb.
- The NO2 sensor SPAN response should be 400 ppb ± 20 ppb (5% span of 400 ppb) with a Std. Dev. < 8 ppb (2% span of 400 ppb). If the sensor response error is greater than ±20 ppb then a GAIN adjustment is required. If the Std. Dev. is greater than 8.0 ppb then the sensor is outside acceptable range and may need replacement.
- The adjusted span response NEW gain should be 0.2 < GAIN < 5.0 and the sensor response 400 ppb ± 20 ppb.

Comments:

Values good, no adjustment needed.

Technician: Jeremy Levine

QA Review: 

MONTROSE AIR QUALITY SERVICES LLC

AEROQUAL AQS-1 NO2 MODULE MULTI-POINT CALIBRATION FORM

Calibration Data on This Form Are For:			Unadjusted Cal.		Adjusted Cal.	X
Network:	City of Detroit (DLZ)	Site:	MTMS Lab		Date:	
Time Off-Line:		Time On-Line:		Technician:	Jeremy Levine	

Calibration Equipment Info.	Analyzer Model:	Aeroqual AQS-1	S/N:	1838	Last Cal:	8/31/23
	Calibrator Model No.:	Teledyne API	S/N:	69	Cal. Date:	3/2/23
	Zero Air Model No.:	Teledyne API	S/N:	n/a	Cert Date:	n/a
	Gas Supplier:	Airgas	Cyl. Cert. Date:	1/26/21	Cyl. Pressure (PSIG)	1,380
	Gas Cylinder ID #:	D068357	Cyl. Conc. (PPM):	30.95	Gas Module Total Flow Rate	130 mL

Analyzer Calibration Settings	"As Found" (Before Any Adjustment)	"As Left" (After Adjustment)
OFFSET	-0.6	
GAIN	1.213	

Calibrator Flow and Test Gas Data					NO ₂ Response		Δ% (Observed Response Vs. Known Conc.) 3	PASS/FAIL
Calibrator Gas Channel		Calibrator Air Channel		Known NO ₂ Gas Conc. (PPB)	Observed from AQS-1			
Display Setting (SLPM)	Actual Flow Rate (SLPM)	Display Setting (SLPM)	Actual Flow Rate (SLPM)			Response (PPB)	Std. Dev. (PPB)	
							#VALUE!	
							#VALUE!	
							#VALUE!	
							#VALUE!	
OFF	OFF	5.0000		0.0			-	

Linear Regression Analysis:					
Slope:	#DIV/0!	Intercept:	#DIV/0!	Corr. Coefficient (r):	#DIV/0!

NOTES:

- The NO₂ sensor zero response should be 0.0 ppb ± 0.2 ppb with a Std. Dev. < 0.2 ppb. If the sensor response error is greater than ± 0.2 ppb then an offset adjustment is required. If the Std. Dev. is greater than 0.2 ppb then the sensor is outside acceptable range and may need replacement.
- The adjusted zero response NEW offset should be -1 < OFFSET < 1 and the sensor response 0.0 ppb ± 0.2 ppb.
- The NO₂ sensor SPAN response should be 400 ppb ± 20 ppb (5% span of 400 ppb) with a Std. Dev. < 8 ppb (2% span of 400 ppb). If the sensor response error is greater than ±20 ppb then a GAIN adjustment is required. If the Std. Dev. is greater than 8.0 ppb then the sensor is outside acceptable range and may need replacement.
- The adjusted span response NEW gain should be 0.2 < GAIN < 5.0 and the sensor response 400 ppb ± 20 ppb.

Comments:

--

Technician: Jeremy Levine

QA Review: 

MONTROSE AIR QUALITY SERVICES LLC

AEROQUAL AQS-1 VOC HIGH RANGE MODULE VERIFICATION/CALIBRATION FORM

Network:	City of Detroit (DLZ)	Site:	MTMS Lab	Date:	9/27/23
Time Off-Line:	10:30 EST	Time On-Line:	12:54 EST	Technician:	Jeremy Levine

Calibration Equipment Info.	Analyzer Model:	Aeroqual AQS-1	S/N:	1838	Last Cal:	9/11/23
	Calibrator Model No:	Teledyne API	S/N:	69	Cal. Date:	3/2/23
	Zero Air Model No:	Teledyne API	S/N:	n/a	Cert Date:	n/a
	Gas Supplier:	AirGas	Cyl. Conc. (PPM):	49.33	Cyl. Pressure (PSIG)	2,080

VOC Sensor Module Calibration Settings	"As Found" (Before Any Adjustment)	"As Left" (After Adjustment)
OFFSET	0.00	
GAIN	1.389	

"AS FOUND" (UNADJUSTED) TEST DATA

Calibrator Flow and Test Gas Data					Observed VOC Response from AQS-1		Error (Δ%)
Calibrator Gas Channel		Calibrator Air Channel		Known VOC Input Gas Conc. (PPM)	Response (PPM)	Std. Dev. (PPM)	
Display Setting (SLPM)	Actual Flow Rate (SLPM)	Display Setting (SLPM)	Actual Flow Rate (SLPM)				
OFF	OFF	5.0000	5.0170	0.00	0.00	0.00	-
0.0500	0.0502	4.9493	4.9735	0.49	0.49	0.00	0.0%
0.0500	0.0501	2.4493	2.4676	0.98	0.92	0.00	-6.1%

"AS LEFT" (ADJUSTED) TEST DATA

Calibrator Flow and Test Gas Data					Observed VOC Response from AQS-1		Error (Δ%)
Calibrator Gas Channel		Calibrator Air Channel		Known VOC Input Gas Conc. (PPM)	Response (PPM)	Std. Dev. (PPM)	
Display Setting (SLPM)	Actual Flow Rate (SLPM)	Display Setting (SLPM)	Actual Flow Rate (SLPM)				
OFF	OFF	5.0000		0.00			-

NOTES:

- The VOC sensor zero response should be 0.0 ppm ± 0.2 ppm with a Std. Dev. < 0.2 ppm. If the sensor response error is greater than ± 0.2 ppm then an offset adjustment is required. If the Std. Dev. is greater than 0.2 ppm then the sensor is outside acceptable range and may need replacement.
- The adjusted zero response NEW offset should be -1 < OFFSET < 1 and the sensor response 0.0 ppm ± 0.2 ppm.
- The VOC sensor SPAN response should be ± 1 ppm (5% span of 20 ppm) with a Std. Dev. < 0.4 ppm (2% span of 20 ppm). If the sensor response error is greater than ± 1 ppm then a GAIN adjustment is required. If the Std. Dev. is greater than 0.4 ppm then the sensor is outside acceptable range and may need replacement.
- The adjusted span response NEW gain should be 0.2 < GAIN < 5.0 and the sensor response 0.0 ppm ± 1 ppm.

Comments:

Values good, no adjustment needed.

Technician: Jeremy Levine

QA Review: 

MONTROSE AIR QUALITY SERVICES LLC

AEROQUAL AQS-1 NO2 MODULE MULTI-POINT CALIBRATION FORM

Calibration Data on This Form Are For:				Unadjusted Cal.	X	Adjusted Cal.	
Network:	City of Detroit (DLZ)	Site:	MTMS Lab	Date:	9/27/23		
Time Off-Line:	12:58 EST	Time On-Line:	16:07 EST	Technician:	Jeremy Levine		

Calibration Equipment Info.	Analyzer Model:	Aeroqual AQS-1	S/N:	1838	Last Cal:	9/11/23
	Calibrator Model No.:	Teledyne API	S/N:	69	Cal. Date:	3/2/23
	Zero Air Model No.:	Teledyne API	S/N:	n/a	Cert Date:	n/a
	Gas Supplier:	Airgas	Cyl. Cert. Date:	1/26/21	Cyl. Pressure (PSIG)	1,300
	Gas Cylinder ID #:	D068357	Cyl. Conc. (PPM):	30.95	Gas Module Total Flow Rate	130 mL

Analyzer Calibration Settings	"As Found" (Before Any Adjustment)	"As Left" (After Adjustment)
OFFSET	-0.6	
GAIN	1.213	

Calibrator Flow and Test Gas Data					NO ₂ Response		Δ% (Observed Response Vs. Known Conc.) 3	PASS/FAIL
Calibrator Gas Channel		Calibrator Air Channel		Known NO ₂ Gas Conc. (PPB)	Observed from AQS-1			
Display Setting (SLPM)	Actual Flow Rate (SLPM)	Display Setting (SLPM)	Actual Flow Rate (SLPM)			Response (PPB)	Std. Dev. (PPB)	
0.0452	0.0453	3.4548	3.4781	397.9	395.0	0.5	-0.7%	
0.0484	0.0485	4.9516	4.9765	298.7	297.3	0.5	-0.5%	
0.0323	0.0324	4.9677	4.9911	199.6	198.1	0.6	-0.8%	
0.0161	0.0163	4.9839	5.0110	100.3	98.7	0.2	-1.6%	
OFF	OFF	5.0000	5.0182	0.0	0.6	0.6	-	

Linear Regression Analysis:					
Slope:	0.993154	Intercept:	0.004498	Corr. Coefficient (r):	0.999991


NOTES:

- The NO₂ sensor zero response should be 0.0 ppb ± 0.2 ppb with a Std. Dev. < 0.2 ppb. If the sensor response error is greater than ± 0.2 ppb then an offset adjustment is required. If the Std. Dev. is greater than 0.2 ppb then the sensor is outside acceptable range and may need replacement.
- The adjusted zero response NEW offset should be -1 < OFFSET < 1 and the sensor response 0.0 ppb ± 0.2 ppb.
- The NO₂ sensor SPAN response should be 400 ppb ± 20 ppb (5% span of 400 ppb) with a Std. Dev. < 8 ppb (2% span of 400 ppb). If the sensor response error is greater than ±20 ppb then a GAIN adjustment is required. If the Std. Dev. is greater than 8.0 ppb then the sensor is outside acceptable range and may need replacement.
- The adjusted span response NEW gain should be 0.2 < GAIN < 5.0 and the sensor response 400 ppb ± 20 ppb.

Comments:

Values good, no adjustment needed.

Technician: Jeremy Levine

QA Review: 

AEROQUAL AQS-1 NO2 MODULE MULTI-POINT CALIBRATION FORM

Calibration Data on This Form Are For:			Unadjusted Cal.		Adjusted Cal.	X
Network:	City of Detroit (DLZ)	Site:	MTMS Lab		Date:	
Time Off-Line:		Time On-Line:		Technician:	Jeremy Levine	

Calibration Equipment Info.	Analyzer Model:	Aeroqual AQS-1	S/N:	1838	Last Cal:	8/31/23
	Calibrator Model No.:	Teledyne API	S/N:	69	Cal. Date:	3/2/23
	Zero Air Model No.:	Teledyne API	S/N:	n/a	Cert Date:	n/a
	Gas Supplier:	Airgas	Cyl. Cert. Date:	1/26/21	Cyl. Pressure (PSIG)	1,380
	Gas Cylinder ID #:	D068357	Cyl. Conc. (PPM):	30.95	Gas Module Total Flow Rate	130 mL

Analyzer Calibration Settings	"As Found" (Before Any Adjustment)	"As Left" (After Adjustment)
OFFSET	-0.6	
GAIN	1.213	

Calibrator Flow and Test Gas Data					NO ₂ Response		Δ% (Observed Response Vs. Known Conc.) 3	PASS/FAIL
Calibrator Gas Channel		Calibrator Air Channel		Known NO ₂ Gas Conc. (PPB)	Observed from AQS-1			
Display Setting (SLPM)	Actual Flow Rate (SLPM)	Display Setting (SLPM)	Actual Flow Rate (SLPM)			Response (PPB)	Std. Dev. (PPB)	
							#VALUE!	
							#VALUE!	
							#VALUE!	
							#VALUE!	
OFF	OFF	5.0000		0.0			-	

Linear Regression Analysis:					
Slope:	#DIV/0!	Intercept:	#DIV/0!	Corr. Coefficient (r):	#DIV/0!

NOTES:

- The NO₂ sensor zero response should be 0.0 ppb ± 0.2 ppb with a Std. Dev. < 0.2 ppb. If the sensor response error is greater than ± 0.2 ppb then an offset adjustment is required. If the Std. Dev. is greater than 0.2 ppb then the sensor is outside acceptable range and may need replacement.
- The adjusted zero response NEW offset should be -1 < OFFSET < 1 and the sensor response 0.0 ppb ± 0.2 ppb.
- The NO₂ sensor SPAN response should be 400 ppb ± 20 ppb (5% span of 400 ppb) with a Std. Dev. < 8 ppb (2% span of 400 ppb). If the sensor response error is greater than ±20 ppb then a GAIN adjustment is required. If the Std. Dev. is greater than 8.0 ppb then the sensor is outside acceptable range and may need replacement.
- The adjusted span response NEW gain should be 0.2 < GAIN < 5.0 and the sensor response 400 ppb ± 20 ppb.

Comments:

--

Technician: Jeremy Levine

QA Review: 

MONTROSE AIR QUALITY SERVICES LLC

AEROQUAL AQS-1 VOC HIGH RANGE MODULE VERIFICATION/CALIBRATION FORM

Network:	City of Detroit (DLZ)	Site:	MTMS Lab	Date:	9/11/23
Time Off-Line:	11:30 EST	Time On-Line:	14:07 EST	Technician:	Jeremy Levine

Calibration Equipment Info.	Analyzer Model:	Aeroqual AQS-1	S/N:	1839	Last Cal:	8/30/23
	Calibrator Model No:	Teledyne API	S/N:	69	Cal. Date:	3/2/23
	Zero Air Model No:	Teledyne API	S/N:	n/a	Cert Date:	n/a
	Gas Supplier:	AirGas	Cyl. Conc. (PPM):	49.33	Cyl. Pressure (PSIG)	2,080

VOC Sensor Module Calibration Settings	"As Found" (Before Any Adjustment)	"As Left" (After Adjustment)
OFFSET	0.00	0.00
GAIN	1.416	1.479

"AS FOUND" (UNADJUSTED) TEST DATA

Calibrator Flow and Test Gas Data					Observed VOC Response from AQS-1		Error (Δ%)
Calibrator Gas Channel		Calibrator Air Channel		Known VOC Input Gas Conc. (PPM)	Response (PPM)	Std. Dev. (PPM)	
Display Setting (SLPM)	Actual Flow Rate (SLPM)	Display Setting (SLPM)	Actual Flow Rate (SLPM)				
OFF	OFF	5.0000	5.0153	0.00	0.00	0.00	-
0.0500	0.0502	4.9493	4.9760	0.49	0.49	0.00	0.0%
0.0500	0.0501	2.4493	2.4680	0.98	0.90	0.00	-8.2%

"AS LEFT" (ADJUSTED) TEST DATA

Calibrator Flow and Test Gas Data					Observed VOC Response from AQS-1		Error (Δ%)
Calibrator Gas Channel		Calibrator Air Channel		Known VOC Input Gas Conc. (PPM)	Response (PPM)	Std. Dev. (PPM)	
Display Setting (SLPM)	Actual Flow Rate (SLPM)	Display Setting (SLPM)	Actual Flow Rate (SLPM)				
OFF	OFF	5.0000	5.0170	0.00	0.00	0.0	-
0.0500	0.0502	4.9493	4.9762	0.49	0.52	0.0	6.1%
0.0500	0.0501	2.4493	2.4686	0.98	0.96	0.0	-2.0%


NOTES:

- The VOC sensor zero response should be 0.0 ppm ± 0.2 ppm with a Std. Dev. < 0.2 ppm. If the sensor response error is greater than ± 0.2 ppm then an offset adjustment is required. If the Std. Dev. is greater than 0.2 ppm then the sensor is outside acceptable range and may need replacement.
- The adjusted zero response NEW offset should be -1 < OFFSET < 1 and the sensor response 0.0 ppm ± 0.2 ppm.
- The VOC sensor SPAN response should be ± 1 ppm (5% span of 20 ppm) with a Std. Dev. < 0.4 ppm (2% span of 20 ppm). If the sensor response error is greater than ± 1 ppm then a GAIN adjustment is required. If the Std. Dev. is greater than 0.4 ppm then the sensor is outside acceptable range and may need replacement.
- The adjusted span response NEW gain should be 0.2 < GAIN < 5.0 and the sensor response 0.0 ppm ± 1 ppm.

Comments:

values good, no adjustment needed.

Technician: Jeremy Levine

QA Review: 

MONTROSE AIR QUALITY SERVICES LLC

AEROQUAL AQS-1 NO2 MODULE MULTI-POINT CALIBRATION FORM

Calibration Data on This Form Are For:				Unadjusted Cal.	X	Adjusted Cal.	
Network:	City of Detroit (DLZ)	Site:	MTMS Lab		Date:	9/11/23	
Time Off-Line:	14:10 EST	Time On-Line:	17:45 EST		Technician:	Jeremy Levine	

Calibration Equipment Info.	Analyzer Model:	Aeroqual AQS-1	S/N:	1839	Last Cal:	8/31/23
	Calibrator Model No.:	Teledyne API	S/N:	69	Cal. Date:	3/2/23
	Zero Air Model No.:	Teledyne API	S/N:	n/a	Cert Date:	n/a
	Gas Supplier:	Airgas	Cyl. Cert. Date:	1/26/21	Cyl. Pressure (PSIG)	1,320
	Gas Cylinder ID #:	D068357	Cyl. Conc. (PPM):	30.95	Gas Module Total Flow Rate	130 mL

Analyzer Calibration Settings	"As Found" (Before Any Adjustment)	"As Left" (After Adjustment)
OFFSET	0.0	
GAIN	1.144	

Calibrator Flow and Test Gas Data					NO ₂ Response		Δ% (Observed Response Vs. Known Conc.) 3	PASS/FAIL
Calibrator Gas Channel		Calibrator Air Channel		Known NO ₂ Gas Conc. (PPB)	<i>Observed from AQS-1</i>			
Display Setting (SLPM)	Actual Flow Rate (SLPM)	Display Setting (SLPM)	Actual Flow Rate (SLPM)			Response (PPB)	Std. Dev. (PPB)	
0.0452	0.0453	3.4548	3.4788	397.8	392.0	0.5	-1.5%	
0.0484	0.0485	4.9516	4.9793	298.6	296.0	0.4	-0.9%	
0.0323	0.0324	4.9677	4.9933	199.5	196.2	0.3	-1.7%	
0.0161	0.0163	4.9839	5.0110	100.3	97.1	0.4	-3.2%	
OFF	OFF	5.0000	5.0171	0.0	-0.8	0.3	-	

Linear Regression Analysis:							
Slope:	0.990537	Intercept:	-1.254498	Corr. Coefficient (r):	0.999979		

NOTES:

AEROQUAL AQS-1 NO2 MODULE MULTI-POINT CALIBRATION FORM

Calibration Data on This Form Are For:				Unadjusted Cal.		Adjusted Cal.	X
Network:	City of Detroit (DLZ)	Site:	MTMS Lab		Date:	8/31/23	
Time Off-Line:		Time On-Line:			Technician:	Jeremy Levine	

Calibration Equipment Info.	Analyzer Model:	Aeroqual AQS-1	S/N:	1839	Last Cal:	8/31/23
	Calibrator Model No.:	Teledyne API	S/N:	69	Cal. Date:	3/2/23
	Zero Air Model No.:	Teledyne API	S/N:	n/a	Cert Date:	n/a
	Gas Supplier:	Airgas	Cyl. Cert. Date:	1/26/21	Cyl. Pressure (PSIG)	1,380
	Gas Cylinder ID #:	D068357	Cyl. Conc. (PPM):	30.95	Gas Module Total Flow Rate	130 mL

Analyzer Calibration Settings	"As Found" (Before Any Adjustment)	"As Left" (After Adjustment)
OFFSET	0.0	
GAIN	1.144	

Calibrator Flow and Test Gas Data					NO ₂ Response		Δ% (Observed Response Vs. Known Conc.) 3	PASS/FAIL
Calibrator Gas Channel		Calibrator Air Channel		Known NO ₂ Gas Conc. (PPB)	<i>Observed from AQS-1</i>			
Display Setting (SLPM)	Actual Flow Rate (SLPM)	Display Setting (SLPM)	Actual Flow Rate (SLPM)			Response (PPB)	Std. Dev. (PPB)	
							#VALUE!	
							#VALUE!	
							#VALUE!	
							#VALUE!	
OFF	OFF	5.0000		0.0			-	

Linear Regression Analysis:					
Slope:	#DIV/0!	Intercept:	#DIV/0!	Corr. Coefficient (r):	#DIV/0!

NOTES:

AEROQUAL AQS-1 VOC HIGH RANGE MODULE VERIFICATION/CALIBRATION FORM

Network:	City of Detroit (DLZ)	Site:	MTMS Lab	Date:	9/27/23
Time Off-Line:	10:30 EST	Time On-Line:	12:54 EST	Technician:	Jeremy Levine

Calibration Equipment Info.	Analyzer Model:	Aeroqual AQS-1	S/N:	1839	Last Cal:	9/11/23
	Calibrator Model No:	Teledyne API	S/N:	69	Cal. Date:	3/2/23
	Zero Air Model No:	Teledyne API	S/N:	n/a	Cert Date:	n/a
	Gas Supplier:	AirGas	Cyl. Conc. (PPM):	49.33	Cyl. Pressure (PSIG)	2,080

VOC Sensor Module Calibration Settings	"As Found" (Before Any Adjustment)	"As Left" (After Adjustment)
OFFSET	0.00	0.00
GAIN	1.479	1.573

"AS FOUND" (UNADJUSTED) TEST DATA

Calibrator Flow and Test Gas Data					Observed VOC Response from AQS-1		Error (Δ%)
Calibrator Gas Channel		Calibrator Air Channel		Known VOC Input Gas Conc. (PPM)	Response (PPM)	Std. Dev. (PPM)	
Display Setting (SLPM)	Actual Flow Rate (SLPM)	Display Setting (SLPM)	Actual Flow Rate (SLPM)				
OFF	OFF	5.0000	5.0170	0.00	0.00	0.00	-
0.0500	0.0502	4.9493	4.9735	0.49	0.47	0.00	-4.1%
0.0500	0.0501	2.4493	2.4676	0.98	0.87	0.00	-11.2%

"AS LEFT" (ADJUSTED) TEST DATA


Calibrator Flow and Test Gas Data					Observed VOC Response from AQS-1		Error (Δ%)
Calibrator Gas Channel		Calibrator Air Channel		Known VOC Input Gas Conc. (PPM)	Response (PPM)	Std. Dev. (PPM)	
Display Setting (SLPM)	Actual Flow Rate (SLPM)	Display Setting (SLPM)	Actual Flow Rate (SLPM)				
OFF	OFF	5.0000	5.0152	0.00	0.00	0.0	-
0.0500	0.0502	4.9493	4.9747	0.49	0.51	0.0	4.1%
0.0500	0.0501	2.4493	2.4696	0.98	0.94	0.0	-4.1%

NOTES:

- The VOC sensor zero response should be 0.0 ppm ± 0.2 ppm with a Std. Dev. < 0.2 ppm. If the sensor response error is greater than ± 0.2 ppm then an offset adjustment is required. If the Std. Dev. is greater than 0.2 ppm then the sensor is outside acceptable range and may need replacement.
- The adjusted zero response NEW offset should be -1 < OFFSET < 1 and the sensor response 0.0 ppm ± 0.2 ppm.
- The VOC sensor SPAN response should be ± 1 ppm (5% span of 20 ppm) with a Std. Dev. < 0.4 ppm (2% span of 20 ppm). If the sensor response error is greater than ± 1 ppm then a GAIN adjustment is required. If the Std. Dev. is greater than 0.4 ppm then the sensor is outside acceptable range and may need replacement.
- The adjusted span response NEW gain should be 0.2 < GAIN < 5.0 and the sensor response 0.0 ppm ± 1 ppm.

Comments:

Technician: Jeremy Levine

QA Review: 

MONTROSE AIR QUALITY SERVICES LLC

AEROQUAL AQS-1 NO2 MODULE MULTI-POINT CALIBRATION FORM

Calibration Data on This Form Are For:				Unadjusted Cal.	X	Adjusted Cal.	
Network:	City of Detroit (DLZ)	Site:	MTMS Lab		Date:	9/27/23	
Time Off-Line:	12:58 EST	Time On-Line:	16:07 EST		Technician:	Jeremy Levine	

Calibration Equipment Info.	Analyzer Model:	Aeroqual AQS-1	S/N:	1839	Last Cal:	9/11/23
	Calibrator Model No.:	Teledyne API	S/N:	69	Cal. Date:	3/2/23
	Zero Air Model No.:	Teledyne API	S/N:	n/a	Cert Date:	n/a
	Gas Supplier:	Airgas	Cyl. Cert. Date:	1/26/21	Cyl. Pressure (PSIG)	1,300
	Gas Cylinder ID #:	D068357	Cyl. Conc. (PPM):	30.95	Gas Module Total Flow Rate	130 mL

Analyzer Calibration Settings	"As Found" (Before Any Adjustment)	"As Left" (After Adjustment)
OFFSET	0.0	
GAIN	1.144	

Calibrator Flow and Test Gas Data					NO ₂ Response		Δ% (Observed Response Vs. Known Conc.) 3	PASS/FAIL
Calibrator Gas Channel		Calibrator Air Channel		Known NO ₂ Gas Conc. (PPB)	<i>Observed from AQS-1</i>			
Display Setting (SLPM)	Actual Flow Rate (SLPM)	Display Setting (SLPM)	Actual Flow Rate (SLPM)		Response (PPB)	Std. Dev. (PPB)		
0.0452	0.0453	3.4548	3.4781	397.9	394.2	0.8	-0.9%	
0.0484	0.0485	4.9516	4.9765	298.7	298.1	0.4	-0.2%	
0.0323	0.0324	4.9677	4.9911	199.6	197.7	0.5	-1.0%	
0.0161	0.0163	4.9839	5.0110	100.3	98.1	0.6	-2.2%	
OFF	OFF	5.0000	5.0182	0.0	-0.2	0.6	-	
Linear Regression Analysis:								
Slope:	0.994563	Intercept:	-0.636495	Corr. Coefficient (r):	0.999975			

NOTES:

AEROQUAL AQS-1 NO2 MODULE MULTI-POINT CALIBRATION FORM

Calibration Data on This Form Are For:				Unadjusted Cal.		Adjusted Cal.	X
Network:	City of Detroit (DLZ)	Site:	MTMS Lab		Date:	8/31/23	
Time Off-Line:		Time On-Line:			Technician:	Jeremy Levine	

Calibration Equipment Info.	Analyzer Model:	Aeroqual AQS-1	S/N:	1839	Last Cal:	8/31/23
	Calibrator Model No.:	Teledyne API	S/N:	69	Cal. Date:	3/2/23
	Zero Air Model No.:	Teledyne API	S/N:	n/a	Cert Date:	n/a
	Gas Supplier:	Airgas	Cyl. Cert. Date:	1/26/21	Cyl. Pressure (PSIG)	1,380
	Gas Cylinder ID #:	D068357	Cyl. Conc. (PPM):	30.95	Gas Module Total Flow Rate	130 mL

Analyzer Calibration Settings	"As Found" (Before Any Adjustment)	"As Left" (After Adjustment)
OFFSET	0.0	
GAIN	1.144	

Calibrator Flow and Test Gas Data					NO ₂ Response		Δ% (Observed Response Vs. Known Conc.) 3	PASS/FAIL
Calibrator Gas Channel		Calibrator Air Channel		Known NO ₂ Gas Conc. (PPB)	<i>Observed from AQS-1</i>			
Display Setting (SLPM)	Actual Flow Rate (SLPM)	Display Setting (SLPM)	Actual Flow Rate (SLPM)			Response (PPB)	Std. Dev. (PPB)	
							#VALUE!	
							#VALUE!	
							#VALUE!	
							#VALUE!	
OFF	OFF	5.0000		0.0			-	

Linear Regression Analysis:					
Slope:	#DIV/0!	Intercept:	#DIV/0!	Corr. Coefficient (r):	#DIV/0!

NOTES:

Using Bios Dry-Cal Flow Standard(s)

APPLICATION INFORMATION:

Calibrator Model/S/N: TAPI T700; SN 69	NETWORK: Marathon Detroit PAMS	SITE: MTMS
Calibration Site: MTMS Site	Test Date: 3/2/2023	
Barometric Pressure (Pa, in mmHg): 743.0	Calibrated by: Jeremy Levine	
Flow Standard Model: Mesa Labs Defender 530+ M, 530+ H	Air Temp. (Ta, in deg. C): 23.1	(=deg. K): 296.3
Flow Standard Base S/N: Not Applicable	Flow Cell Model No: Defender 530+ M Defender 530+ H	
Certification Date: Not Applicable	Flow Cell S/N: 205428	205361
	Flow Cell Certification Date: 7/22/2022	7/21/2022

Check One:

Air Channel

Gas Channel

(X) MFC Drive Voltage (mVDC)	Flow Meter Readings (5 sets of 10 averaged flows)					Average Flow (F1...F5) (SLPM)	STD DEV F1...F5 (in <i>sccm</i>)	Flow Rate <i>From Previous</i> <u>Cal</u> (SLPM)	$\Delta\%$ ("New Cal Flow" Vs "Prev. Cal Flow")
	F ₁ (SLPM)	F ₂ (SLPM)	F ₃ (SLPM)	F ₄ (SLPM)	F ₅ (SLPM)				
5000	10.8520	10.8580	10.8420	10.8600	10.8530	10.853	7.0	10.832	-0.2%
4750	10.3170	10.3160	10.3150	10.3110	10.3160	10.315	2.3	10.266	-0.5%
4500	9.6906	9.6895	9.6869	9.6877	9.6923	9.689	2.2	9.708	0.2%
4250	9.1475	9.1495	9.1520	9.1448	9.1438	9.148	3.4	9.157	0.1%
4000	8.5987	8.6045	8.6008	8.6002	8.5984	8.601	2.4	8.603	0.0%
3750	8.0527	8.0573	8.0563	8.0529	8.0549	8.055	2.0	8.053	0.0%
3500	7.5167	7.5172	7.5134	7.5132	7.5105	7.514	2.8	7.507	-0.1%
3250	6.9823	6.9845	6.9790	6.9783	6.9767	6.980	3.2	6.967	-0.2%
3000	6.4503	6.4485	6.4492	6.4473	6.4441	6.448	2.4	6.430	-0.3%
2750	5.9049	5.8928	5.8966	5.9052	5.9054	5.901	5.9	5.879	-0.4%
2500	5.3137	5.3172	5.3185	5.3195	5.3172	5.317	2.2	5.334	0.3%
2250	4.7718	4.7757	4.7813	4.7790	4.7793	4.777	3.7	4.801	0.5%
2000	4.2360	4.2314	4.2315	4.2332	4.2360	4.234	2.3	4.265	0.7%
1750	3.6825	3.6817	3.6854	3.6793	3.6879	3.683	3.3	3.724	1.1%
1500	3.1393	3.1393	3.1519	3.1439	3.1461	3.144	5.3	3.189	1.4%
1250	2.6238	2.6284	2.6290	2.6273	2.6287	2.627	2.1	2.650	0.9%
1000	2.0926	2.0917	2.0912	2.0918	2.0917	2.092	0.5	2.115	1.1%
750	1.5499	1.5498	1.5498	1.5505	1.5516	1.550	0.8	1.579	1.8%
500	1.0163	1.0157	1.0146	1.0148	1.0145	1.015	0.8	1.037	2.1%
250	0.48024	0.48137	0.48059	0.48179	0.48179	0.481	0.7	0.493	2.4%
SLOPE:	0.002180501		INTERCEPT: -0.102560589		CORRELATION COEFF (r):		0.999962608		

Comments:

Technician:

Jeremy Levine

3/2/2023

(signature)

Date

TAPI T700 MFC Calibration Using Bios Dry-Cal Flow Standard(s)

CALIBRATOR APPLICATION INFORMATION:

Calibrator Model/S/N:	TAPI T700; SN 69	NETWORK:	Marathon Detroit PAMS	SITE:	MTMS	
Calibration Site:	MTMS Site	Test Date:	3/1/2023			
Barometric Pressure (Pa, in mmHg):	740.0	Calibrated by:	J Levine			
Flow Standard Model:	Mesa Labs Defender 530+ L	Air Temp. (Ta, in deg. C):	22.9	(=deg. K):	296.1	
Flow Standard Base S/N:	Not Applicable	Flow Cell Model No:	530+ Low Flow			
Base Certification Date:	Not Applicable	Flow Cell S/N:	205663			
		Flow Cell Certification Date:	8/4/2022			

Check One: Air Channel **X** Gas Channel

(X) MFC Drive Voltage (mVDC)	0.0538 (5 sets of 10 averaged flows)					Average Flow (F1...F5) (SLPM)	STD DEV F1...F5 (in <u>sccm</u>)	Flow Rate <i>From Previous</i> Cal (SLPM)	Δ% ("New Cal Flow" Vs "Prev. Cal Flow")
	F ₁ (SLPM)	F ₂ (SLPM)	F ₃ (SLPM)	F ₄ (SLPM)	F ₅ (SLPM)				
5000	0.05511	0.05513	0.05514	0.05507	0.05504	0.0551	0.04	0.0549	-0.3%
4750	0.05239	0.05240	0.05240	0.05241	0.05240	0.0524	0.01	0.0523	-0.2%
4500	0.04958	0.04960	0.04961	0.04963	0.04964	0.0496	0.02	0.0496	-0.1%
4250	0.04669	0.04674	0.04676	0.04680	0.04683	0.0468	0.05	0.0468	0.1%
4000	0.04415	0.04416	0.04412	0.04406	0.04405	0.0441	0.05	0.0441	0.1%
3750	0.04147	0.04148	0.04146	0.04150	0.04145	0.0415	0.02	0.0414	-0.2%
3500	0.03870	0.03873	0.03874	0.03873	0.03875	0.0387	0.02	0.0387	-0.2%
3250	0.03587	0.03589	0.03591	0.03593	0.03597	0.0359	0.04	0.0359	-0.1%
3000	0.03329	0.03327	0.03324	0.03320	0.03318	0.0332	0.05	0.0331	-0.5%
2750	0.03054	0.03055	0.03056	0.03057	0.03056	0.0306	0.01	0.0304	-0.5%
2500	0.02775	0.02777	0.02778	0.02779	0.02781	0.0278	0.02	0.0277	-0.5%
2250	0.02502	0.02498	0.02500	0.02498	0.02499	0.0250	0.02	0.0249	-0.3%
2000	0.02231	0.02232	0.02232	0.02232	0.02230	0.0223	0.01	0.0222	-0.7%
1750	0.01950	0.01951	0.01952	0.01952	0.01953	0.0195	0.01	0.0193	-0.9%
1500	0.01668	0.01667	0.01667	0.01668	0.01669	0.0167	0.01	0.0166	-0.6%
1250	0.01392	0.01393	0.01394	0.01394	0.01392	0.0139	0.01	0.0138	-0.9%
1000	0.01110	0.01106	0.01107	0.01110	0.01111	0.0111	0.02	0.0110	-1.0%
750	0.00831	0.00831	0.00832	0.00832	0.00831	0.0083	0.01	0.0082	-1.3%
500	0.00548	0.00545	0.00547	0.00546	0.00546	0.0055	0.01	0.0054	-1.9%
250	0.00261	0.00264	0.00262	0.00262	0.00262	0.0026	0.01	0.0025	-6.0%
SLOPE: 0.000011			INTERCEPT: 0.000167747			CORRELATION COEFF (r): 0.999979775			

Comments: _____

Technician: Jeremy Levine

(signature)

3/1/23

Date

CERTIFICATE OF ANALYSIS

Grade of Product: TRACEABILITY STANDARD

Part Number: X02NI99T33W0004	Reference Number: 54-402006473-1
Cylinder Number: D068357	Cylinder Volume: 32.0 CF
Laboratory: 124 - Chicago (SAP) - IL	Cylinder Pressure: 2218 PSIG
	Valve Outlet: 660
	Certification Date: Jan 26, 2021

Expiration Date: Jan 26, 2024

This cylinder has been analytically certified as directly traceable to NIST with a total analytical uncertainty as stated below with a confidence level of 95%, in accordance with Airgas ISO procedures. There are no significant impurities which affect the use of this calibration mixture. All concentrations are on a mole/mole basis unless otherwise noted.


Do Not Use This Cylinder Below 100 psig.

ANALYTICAL RESULTS					
Component	Requested Concentration	Actual Concentration	Total Relative Uncertainty		
NITROGEN DIOXIDE	30.00 PPM	30.95 PPM	+/- 1% NIST Traceable		
NITROGEN	Balance				
CALIBRATION STANDARDS					
Type	Lot ID	Cylinder No	Concentration	Uncertainty	Expiration Date
GMIS	401438584104	EB0120492	48.18 PPM NITROGEN DIOXIDE/NITROGEN	+/- 1.8%	Nov 01, 2022
ANALYTICAL EQUIPMENT					
Instrument/Make/Model	Analytical Principle			Last Multipoint Calibration	
MKS FTIR NO2 017707558	FTIR			Jan 07, 2021	

Triad Data Available Upon Request

PERMANENT NOTES: OXYGEN ADDED TO MAINTAIN STABILITY




 Approved for Release

CERTIFICATE OF ANALYSIS

Grade of Product: CERTIFIED STANDARD-SPEC

Part Number:	X02NI99C15A0104	Reference Number:	141-402072346-1
Cylinder Number:	EB0112566	Cylinder Volume:	144.4 CF
Laboratory:	124 - Stryker (SAP) - OH	Cylinder Pressure:	2015 PSIG
Analysis Date:	Mar 31, 2021	Valve Outlet:	350
Lot Number:	141-402072346-1		

Product composition verified by direct comparison to calibration standards traceable to N.I.S.T. weights and/or N.I.S.T. Gas Mixture reference materials.

ANALYTICAL RESULTS

Component	Req Conc	Actual Concentration (Mole %)	Analytical Uncertainty
ISOBUTYLENE	50.00 PPM	49.33 PPM	+/- 2%
NITROGEN	Balance		





Approved for Release


Appendix B: Calibration Certification Sheets

GENERAL INFORMATION

Instrument Type	Dust Sentry Pro
Serial Number	DP 11102021-1838

Aeroqual Connect

Version	V1.18.0	OS Image	V4.1.18.0
WiFi SSID	DP 11102021-1838	Password	Aeroqual
Default User	Administrator	Password	aqmadmin
Sensor List	DP_SensorList_V8.3.6.aql		

 Please contact Aeroqual for login and password to access your instrument on Aeroqual Cloud (<http://cloud.aeroqual.com>).

Instrument Configuration

Particle Channels		Environmental Channels		Communication / Software
<input checked="" type="checkbox"/>	TSP	8PC 0.3	TEMP	<input checked="" type="checkbox"/> Connect
<input checked="" type="checkbox"/>	PM 1	8PC 0.5	RH	Support
<input checked="" type="checkbox"/>	PM 2.5	8PC 0.7	ITEMP	Basic
<input checked="" type="checkbox"/>	PM 10	8PC 1.0	WS	Plus
		8PC 2.0	WD	3G modem
		8PC 2.5	AN1	
		8PC 3.0	AN2	
		8PC 5.0	AN3	
		8PC 10	Freq	
			ALT	
			Pyrano	
			Leq	

Integrated Modules

Type	Serial No.	QC	Type	Serial No.	QC
ARK1124C	KSA5124761	Pass	Met One 9722-1	B14058	Pass
Pump Module	AQM PMP03 2804211-039	Pass	Electronics Module	AQM M11O 2108101-042	

 For technical, maintenance and service information, please refer to Dust Monitor User Guide or contact Aeroqual for access to free online training (<http://training.aeroqual.com>).

PERFORMANCE REPORT

Calibration Data

Item	Value	Unit	Item	Value	Unit
Sample System Leak Tightness	Pass	/	Inlet Heater	Pass	/
Sample Flow Rate	1.004	SLPM	Sheath Flow Rate	1.301	SLPM
Zero Filter Reference Reading	0.00	µg/m ³			

Standards Used

Standard	Make	Serial Number	Calibration Due
Vacuum Gauge	SMC	VAC005	N/A
Flow Meter	TSI	4140 1438 025	24/11/2021

Activate Negative Number Filters on all gas and dust channels: YES

FACTORY MODULE SETTINGS

MOD ULE	VER SION	H0	H1	H2	H3	TIMA	TIMR	TEMA	TEMR	PWML	PWMH	HTR	GAIN	Gain	Offset
PM1	4.0.0	0.000	1.000	760.000	298.150	2	60	1.4	2	0	0	1.00	1	1.000	0.000
PM2.5	4.0.0	0.000	1.000	760.000	298.150	2	60	1.4	2	0	0	1.00	1	1.000	0.000
PM10	4.0.0	0.000	1.000	760.000	298.150	2	60	1.4	2	0	0	1.00	1	1.000	0.000
TSP	4.0.0	0.000	1.000	760.000	298.150	2	60	1.4	2	0	0	1.00	1	1.000	0.000
Battery voltage	1.3.0	7.200	0.000	0.000	0.000	0	1000	0	0	0	0	0.00	0	0.000	0.000

Approvals

QC Technician	@ Ariyan Hassan	QC Approval	@ Jeremy Turner
Date	12 Oct 2021	Date	13 Oct 2021

Calibration Certificate(s)

Met One
Instruments1600 Washington Blvd
Grants Pass, OR 97526
(541) 471-7111
(541) 471-7118 (Fax)
Service@metone.com

Calibration Certificate

The calibration results on this report certify that this instrument complies with the product specifications at the time of calibration. Calibration was performed according to accepted industry methods using equipment, procedures, and standards that are traceable to NIST and ISO.

Recommended calibration interval is 12 months from the first day of use.

Instrument Model# 9722-1Instrument Serial# B14058Date of Calibration 5/7/2021Sensor # 1362

Brittney Wentowski
Calibration Technician

R. Jones
Quality Check

Temperature 24 °CRelative Humidity 26 %Test Procedure: 9722-6100

PSL Size (µm)	Test Results	Test Spec.	Lot# NIST	Expiration
0.3	Pass	± 10%	223077	04/30/2023
0.5	Pass	± 10%	219480	11/30/2022
0.7	Pass	± 10%	229561	08/31/2023
1.0	Pass	± 10%	229294	8/31/2023
2.0	Pass	± 10%	231222	09/30/2023
3.0	Pass	± 10%	231458	09/30/2023
5.0	Pass	± 10%	214115	07/31/2022
10.0	Pass	± 10%	230028	09/30/2023

Standards	Model	SN	Cal Due
Particle Counter	GT-526S	X17420	6/20/2021
DMM	117 Multimeter	49320156	6/15/2021
FLOWMETER	4040	40401945009	1/13/2022
RH/TEMP SENSOR	G3120	G4587	2/2/2022

This calibration certificate shall not be reproduced except in full, without the written approval of Met One Instruments Inc.

TEST REPORT

Product family WXT530 series
Product type WXT536
Order code 6B1B2A2D1A1B
Serial number U1270032
Manufacturer Vaisala Oyj, Finland
Test date 27 March 2022

This test report certifies that the product was thoroughly tested and inspected, and found to meet its published test limits when it was shipped from Vaisala.

Test results

Test	Result	Lower limit	Upper limit	Unit
Rain response	385	345	575	mV
Zero wind speed	0	0	0.4	m/s
Pressure difference	0.09	-1	1	hPa
Temperature difference	-0.87	-2	2	°C
Humidity difference	-1.44	-10	10	%RH
Heating current	0.73	0.6	0.8	A
Current (service port)	4.23	0.5	6	mA
Communication (service port)	pass	PASS	PASS	-
Current (main port)	3.68	0.5	6	mA
Communication (main port)	pass	PASS	PASS	-

Ambient conditions / Humidity 8.65 ±5 %RH, Temperature 22.05 ±1 °C, Pressure 1022.07 ±1 hPa.

Signature

Technician



Instrument: PTUMODULE
Serial Number: U1130038
Manufacturer: Vaisala Oyj
Issue Date: 2022-03-27

Approved by:

Digitally signed by: Saastamoinen Anssi
 Date: 2022-03-27 10:53:11 (+03:00)
 Location: Vaisala Oyj, Finland

The humidity sensor of the instrument was calibrated by comparing the instrument's humidity reading to a generated reference humidity reading. The reference humidity reading was calculated based on two-pressure humidity generation principle, using the measurement results of saturator pressure and temperature and calibration chamber pressure and temperature.

The temperature sensor of the instrument was calibrated by comparing the instrument's temperature readings to a reference thermometer.

The pressure sensor of the instrument was calibrated by comparing the instrument's pressure readings to a reference barometer.

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k = 2$, which for a normal distribution corresponds to a coverage probability of approximately 95 %. The measurement results are traceable to the international system of units (SI) through national metrology institutes (NIST USA, MIKES Finland, or equivalent) or via ISO/IEC 17025 accredited calibration laboratories.

Humidity and temperature calibration results, calibration date 2022-03-23

Reference Humidity [%rh]	Reference Temperature [°C]	Observed Humidity [%rh]	Observed Temperature [°C]	Humidity Error [%rh]	Acceptance Limit [%rh]
0.1	22.52	0.0	22.51	-0.1	±3.0
15.0	22.52	14.7	22.51	-0.3	±3.0
33.0	22.52	32.8	22.52	-0.2	±3.0
54.0	22.53	53.9	22.53	-0.1	±3.0
75.1	22.53	75.1	22.53	0.0	±3.0
95.3	22.54	96.4	22.54	1.1	±5.0

Reference Temperature [°C]	Observed Temperature [°C]	Temperature Error [°C]	Acceptance Limit [°C]
22.53	22.53	0.00	±0.30

Ambient conditions in humidity and temperature calibration

Humidity [%rh] Temperature [°C] Pressure [hPa]
 19 ±4 25 ±2 1019 ±20

Reference equipment used in Humidity and temperature calibration

Type	Identity Number	Certificate Number	Calibration date	Calibration due date
PTU307	19542	K008-E05564	2021-11-08	2022-11-30
PXI Pt-100 sensor	19923	K008-E06355	2021-12-12	2022-12-31
DPS823B	19906	K008-E05706	2021-11-15	2022-11-30
PXI Pt-100 sensor	19921	K008-E06357	2021-12-12	2022-12-31
PXIe-4080	19920	E06358	2021-12-13	2022-12-31

Pressure calibration results, calibration date 2022-03-16

Reference Pressure [hPa]	Observed Pressure [hPa]	Pressure Error [hPa]	Acceptance limit [hPa]
601.3	601.3	0.0	±0.5
800.6	800.6	0.0	±0.5
901.0	901.0	0.0	±0.5
1080.8	1080.8	0.0	±0.5

Reference equipment used in pressure calibration

Type	Identity Number	Certificate Number	Calibration date	Calibration due date
Fluke RPM4	17966	E06297	2021-12-09	2022-06

Calibration uncertainty (k=2, ~95% confidence level):

Humidity ±0.6 %rh @ 0...40 %rh, ±1.0 %rh @ 40...95 %rh
 Temperature ±0.10 °C
 Pressure ±0.3 hPa

GENERAL INFORMATION

Instrument Type	Dust Sentry Pro
Serial Number	DP 11102021-1839

Aeroqual Connect

Version	V1.18.0	OS Image	V4.1.18.0
WiFi SSID	DP 11102021-1839	Password	Aeroqual
Default User	Administrator	Password	aqmadmin
Sensor List	DP_SensorList_V8.3.6.aql		

Please contact Aeroqual for login and password to access your instrument on Aeroqual Cloud (<http://cloud.aeroqual.com>).

Instrument Configuration

Particle Channels		Environmental Channels		Communication / Software
<input checked="" type="checkbox"/> TSP	8PC 0.3	<input type="checkbox"/> TEMP	<input type="checkbox"/> RAIN	<input checked="" type="checkbox"/> Connect
<input checked="" type="checkbox"/> PM 1	8PC 0.5	<input type="checkbox"/> RH	<input type="checkbox"/> SOLAR	Support
<input checked="" type="checkbox"/> PM 2.5	8PC 0.7	<input type="checkbox"/> ITEMP	<input type="checkbox"/> HAIL	Basic
<input checked="" type="checkbox"/> PM 10	8PC 1.0	<input type="checkbox"/> WS	<input type="checkbox"/> PRESS	Plus
	8PC 2.0	<input type="checkbox"/> WD	<input type="checkbox"/> AIR T	3G modem
	8PC 2.5	<input type="checkbox"/> AN1	<input type="checkbox"/> AIR RH	
	8PC 3.0	<input type="checkbox"/> AN2	<input type="checkbox"/> LAT	
	8PC 5.0	<input type="checkbox"/> AN3	<input type="checkbox"/> LON	
	8PC 10	<input type="checkbox"/> Freq	<input type="checkbox"/> ALT	
			<input type="checkbox"/> Pyrano	
			<input type="checkbox"/> Leq	

Integrated Modules

Type	Serial No.	QC	Type	Serial No.	QC
ARK1124C	KSA5124614	Pass	Met One 9722-1	B14186	Pass
Pump Module	AQM PMP03 2804211-023	Pass	Electronics Module	AQM M1IO 2108101-074	

For technical, maintenance and service information, please refer to Dust Monitor User Guide or contact Aeroqual for access to free online training (<http://training.aeroqual.com>).

PERFORMANCE REPORT

Calibration Data

Item	Value	Unit	Item	Value	Unit
Sample System Leak Tightness	Pass	/	Inlet Heater	Pass	/
Sample Flow Rate	0.999	SLPM	Sheath Flow Rate	1.251	SLPM
Zero Filter Reference Reading	0.00	µg/m ³			

Standards Used

Standard	Make	Serial Number	Calibration Due
Vacuum Gauge	SMC	VAC005	N/A
Flow Meter	TSI	4140 1438 025	24/11/2021

Activate Negative Number Filters on all gas and dust channels: YES

FACTORY MODULE SETTINGS

MOD ULE	VER SION	H0	H1	H2	H3	TIMA	TIMR	TEMA	TEMR	PWML	PWMH	HTR	GAIN	Gain	Offset
PM1	4.0.0	0.000	1.000	760.000	298.150	2	60	1.4	2	0	0	1.00	1	1.000	0.000
PM2.5	4.0.0	0.000	1.000	760.000	298.150	2	60	1.4	2	0	0	1.00	1	1.000	0.000
PM10	4.0.0	0.000	1.000	760.000	298.150	2	60	1.4	2	0	0	1.00	1	1.000	0.000
TSP	4.0.0	0.000	1.000	760.000	298.150	2	60	1.4	2	0	0	1.00	1	1.000	0.000
Battery voltage	1.3.0	7.200	0.000	0.000	0.000	0	1000	0	0	0	0	0.00	0	0.000	0.000

Approvals

QC Technician	@ Ariyan Hassan	QC Approval	@ Jeremy Turner
Date	12 Oct 2021	Date	13 Oct 2021

Calibration Certificate(s)



Met One
Instruments


1600 Washington Blvd
Grants Pass, OR 97526
(541) 471-7111
(541) 471-7116 (Fax)
Service@metone.com

Calibration Certificate

The calibration results on this report certify that this instrument complies with the product specifications at the time of calibration. Calibration was performed according to accepted industry methods using equipment, procedures, and standards that are traceable to NIST and ISO.

Recommended calibration interval is 12 months from the first day of use.

Instrument Model# 9722-1 Instrument Serial# B14186
Date of Calibration 5/19/2021 Sensor # 1379

Brittney Wentowski 
Calibration Technician

Calibration Technician

Quality Check

Temperature 24 °C

Relative Humidity 26 %

Test Procedure: 9722-6100

PSL Size (µm)	Test Results	Test Spec.	Lot# NIST	Expiration
0.3	Pass	± 10%	223077	04/30/2023
0.5	Pass	± 10%	219480	11/30/2022
0.7	Pass	± 10%	229561	08/31/2023
1.0	Pass	± 10%	229294	8/31/2023
2.0	Pass	± 10%	231222	09/30/2023
3.0	Pass	± 10%	231458	09/30/2023
5.0	Pass	± 10%	214115	07/31/2022
10.0	Pass	± 10%	230028	09/30/2023

Standards	Model	SN	Cal Due
Particle Counter	GT-526S	X17420	6/20/2021
DMM	117 Multimeter	49320156	6/15/2021
FLOWMETER	4040	40401945009	1/13/2022
RH/TEMP SENSOR	G3120	G4587	2/2/2022

This calibration certificate shall not be reproduced except in full, without the written approval of Met One Instruments Inc.



Instrument: PTUMODULE
Serial Number: U1130042
Manufacturer: Vaisala Oyj
Issue Date: 2022-03-27

Approved by:

Digitally signed by: Gaastamoinen Anssi
 Date: 2022-03-27 11:00:15 (+03:00)
 Location: Vaisala Oyj, Finland

The humidity sensor of the instrument was calibrated by comparing the instrument's humidity reading to a generated reference humidity reading. The reference humidity reading was calculated based on two-pressure humidity generation principle, using the measurement results of saturator pressure and temperature and calibration chamber pressure and temperature.

The temperature sensor of the instrument was calibrated by comparing the instrument's temperature readings to a reference thermometer.

The pressure sensor of the instrument was calibrated by comparing the instrument's pressure readings to a reference barometer.

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k = 2$, which for a normal distribution corresponds to a coverage probability of approximately 95%. The measurement results are traceable to the international system of units (SI) through national metrology institutes (NIST USA, MIKES Finland, or equivalent) or via ISO/IEC 17025 accredited calibration laboratories.

Humidity and temperature calibration results, calibration date 2022-03-23

Reference Humidity [%rh]	Reference Temperature [°C]	Observed Humidity [%rh]	Observed Temperature [°C]	Humidity Error [%rh]	Acceptance Limit [%rh]
0.1	22.52	0.0	22.52	-0.1	±3.0
15.0	22.52	14.6	22.53	-0.4	±3.0
33.0	22.52	32.6	22.53	-0.4	±3.0
54.0	22.53	53.8	22.53	-0.2	±3.0
75.1	22.53	75.1	22.53	0.0	±3.0
95.3	22.54	96.5	22.53	1.2	±5.0

Reference Temperature [°C]	Observed Temperature [°C]	Temperature Error [°C]	Acceptance Limit [°C]
22.53	22.53	0.00	±0.30

Ambient conditions in humidity and temperature calibration

Humidity [%rh] Temperature [°C] Pressure [hPa]
 19 ±4 25 ±2 1019 ±20

Reference equipment used in Humidity and temperature calibration

Type	Identity Number	Certificate Number	Calibration date	Calibration due date
PTU307	19542	K008-E05564	2021-11-08	2022-11-30
PXI Pt-100 sensor	19923	K008-E06355	2021-12-12	2022-12-31
DPS823B	19906	K008-E05706	2021-11-15	2022-11-30
PXI Pt-100 sensor	19921	K008-E06357	2021-12-12	2022-12-31
PXIe-4080	19920	E06358	2021-12-13	2022-12-31

Pressure calibration results, calibration date 2022-03-16

Reference Pressure [hPa]	Observed Pressure [hPa]	Pressure Error [hPa]	Acceptance limit [hPa]
601.3	601.3	0.0	±0.5
800.6	800.6	0.0	±0.5
901.0	901.0	0.0	±0.5
1080.8	1080.8	0.0	±0.5

Reference equipment used in pressure calibration

Type	Identity Number	Certificate Number	Calibration date	Calibration due date
Fluke RPM4	17966	E06297	2021-12-09	2022-06

Calibration uncertainty (k=2, ~95% confidence level):

Humidity ±0.6 %rh @ 0...40 %rh, ±1.0 %rh @ 40...95 %rh
 Temperature ±0.10 °C
 Pressure ±0.3 hPa

TEST REPORT

Product family WXT530 series
Product type WXT536
Order code 6B1B2A2D1A1B
Serial number U1270033
Manufacturer Vaisala Oyj, Finland
Test date 27 March 2022

This test report certifies that the product was thoroughly tested and inspected, and found to meet its published test limits when it was shipped from Vaisala.

Test results

Test	Result	Lower limit	Upper limit	Unit
Rain response	381	345	575	mV
Zero wind speed	0	0	0.4	m/s
Pressure difference	0.06	-1	1	hPa
Temperature difference	-0.81	-2	2	°C
Humidity difference	-1.43	-10	10	%RH
Heating current	0.74	0.6	0.8	A
Current (service port)	4.02	0.5	6	mA
Communication (service port)	pass	PASS	PASS	-
Current (main port)	3.41	0.5	6	mA
Communication (main port)	pass	PASS	PASS	-

Ambient conditions / Humidity 8.67 ±5 %RH, Temperature 22.09 ±1 °C, Pressure 1022.06 ±1 hPa.

Signature

Technician

Appendix C: Locations of MI EGLE Monitors Relative to the Former State Fairgrounds

Locations of MI EGLE Monitors Relative to the Former State Fairgrounds

