

April 11, 2016

ATC Associates
46555 Humboldt Dr.
Suite 100
Novi, MI 48377

Subject: Bates Academy
188BS16170

Dear Mr. Smith :

Thank you for making Brighton Analytical, L.L.C. your laboratory of choice. Attached are the results for the samples submitted on 04/04/2016 for the above mentioned project. NELAP/TNI Accredited Analysis and MDEQ Drinking Water Certified Analysis will be identified in their respective reporting formats. Hard copies can be supplied at your request for a fee of \$20.00 per copy.

The invoice for this project will be emailed separately. If you have any questions concerning the data or invoice, please don't hesitate to contact our office. We welcome your comments and suggestions to improve our quality systems. Please reference Brighton Analytical, L.L.C. Project ID 38360 when calling or emailing. We thank you for this opportunity to partner with you on this project and hope to work with you again in the future.

Sincerely,
Brighton Analytical, L.L.C.

BA

Brighton Analytical, L.L.C.
 Email: bai-brighton@shcglobal.net
 2105 Pless Drive
 Brighton, MI 48114
 Phone: 810-229-7575
 Fax: 810-229-8650

PROJECT NAME: BATES Academy

PROJECT #: 188135 16170

PO #: (PLEASE NOTE IF DIFFERENT BILLING ADDRESS)

BA PROJECT #: 3840

ABBREVIATIONS FOR MATRIX

S = Solid
L = Liquid
DW = Drinking H₂O
O = Oil
P = Wipe
A = Air (Teller Bag)
F = Filter
T = Tube
M = Misc.

Sample collected by: Dawn Wintire

REQUESTED TURNAROUND: (circle one)
 Rush: 1-3 business days (verify with lab & specify date needed)
 1 Day = 25X Cost 2 Day = 2X Cost 3 Day = 1.5X Cost
 Standard: 5 business days

Brighton ID #	Sample Description	Container Type & Quantity				Sample Matrix				Analysis Requested/Method				COMPANY/MAILING ADDRESS:
		VOA'S (PRES) Y N N/A	HDPE UNPRESERVED	HDPE HNO ₃	HDPE H ₂ SO ₄	HDPE NAOH	AMBER PRESERVED?	GLASS, NO PRESERVATIVE	STERILIZED BACTERIA	MEOH Preserved Y N	LEAD - Primary (P)	COPPER - Primary (P)	LEAD - FLUSH (F) - HOLD	COPPER - FLUSH (F) - HOLD
1) 1-DW-F-P-Bates pre-k #10	4/1 16:15										X	X		
2) 1-DW-F-P-Bates pre-k #10	4/1 16:10										X	X		
3) 1-DW-F-P-Bates pre-k #10	4/1 16:13										X	X		
4) 1-DW-F-P-Bates pre-k #10	4/1 16:08										X	X		
5) 3-K-P-Bates - K-K-HW	4/1 16:24										X	X		
6) 3-K-P-Bates - K-K-HW	4/1 16:24										X	X		
7)														
8)														
9)														
10)														

Special Instructions: If Lead or Copper is above detection limits - prioritize Full Sample

Please fill out the Chain of Custody completely and review. Incorrect or incomplete information will result in a "hold" on all analyses.

Trans. #	RELINQUISHED BY:	RECEIVED BY:	DATE:	TIME:	Trans. #	RELINQUISHED BY:	RECEIVED BY:	DATE:	TIME:
1	<u>Dawn Wintire</u>	<u>Dawn Wintire</u>	<u>4-11-16</u>	<u>12:32pm</u>	3				
2					4				

PAGE 1 OF 1

COMPANY/MAILING ADDRESS:

ATC Group Services, LLC
46555 Hawthorn Blvd DE STE 100
NEWT, ME 04857
 ATTN: Robert Smith
 PHONE: (207) 669-5140
 FAX OR EMAIL: Robert.Smith@atcservices.com

Samples received within hold time? yes ☒ no ☐

Temperature of samples cont. Min

pHs verified in login? yes ☒ no ☐

Headspace/bubbles in VOA's? yes ☐ no ☒ n/a ☒

Sample containers and COC match? yes ☒ no ☐

BILLING ADDRESS (IF REQUIRED):

Same as mailing

* Not on container

Drinking H₂O:

Fax to LCHD? yes ☐ no ☒

Chlorinated Water Supply? yes ☒ no ☐

AMT.: 100

MCL Failure: yes ☐ no ☒

Client Notified (date/time/initials): _____



Brighton Analytical LLC
2105 Pless Drive
Brighton, Michigan 48114
Phone: (810)229-7575 (810)229-8650
e-mail: bai-brighton@sbcglobal.net
MDNRE Certified #9404
NELAC Accredited #176507

Sample Date/Time: 4/1/2016 16:15
Submit Date/Time: 4/4/2016 12:32
Report Date: 4/11/2016

ATC Associates
46555 Humboldt Dr.
Suite 100
Novi, MI 48377

BA Project # **38360**
BA Sample ID **CC09107**


Project Name: **Bates Academy**
Project Number: **188BS16170**
Sample ID: **1-DWF-P-Bates Prek #10**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
Drinking Water Metal Analysis							
Total Copper (Drinking Water)	350	ug/L	20	1300	EPA 200.8 rev5.4	19:40	04/08/2016
Total Lead (Drinking Water)	4	ug/L	1	15	EPA 200.8 rev5.4	19:40	04/08/2016

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 
Date 4/11/16



Brighton Analytical LLC
2105 Pless Drive
Brighton, Michigan 48114
Phone: (810)229-7575 (810)229-8650
e-mail: bai-brighton@sbcglobal.net
MDNRE Certified #9404
NELAC Accredited #176507

Sample Date/Time: 4/1/2016 16:13
Submit Date/Time: 4/4/2016 12:32
Report Date: 4/11/2016

ATC Associates
46555 Humboldt Dr.
Suite 100
Novi, MI 48377

BA Project # **38360**
BA Sample ID **CC09109**

Project Name: **Bates Academy**
Project Number: **188BS16170**
Sample ID: **2-DWF-P-Bates-Hall-by Gym**

Analyte Name	Result	Units	RL	MCL	Method Reference	Analysis Time	Analysis Date
Drinking Water Metal Analysis							
Total Copper (Drinking Water)	30	ug/L	20	1300	EPA 200.8 rev5.4	19:44	04/08/2016
Total Lead (Drinking Water)	Not detected	ug/L	1	15	EPA 200.8 rev5.4	19:44	04/08/2016

RL=Reported detection limit for analytical method requested. Some compounds require special analytical methods to achieve MDNR designated target detection limits (TDL).

MCL = Maximum contaminant Levels.

Analysis not specifically identified as drinking water are for non-regulatory compliance purposes.

Released by 
Date 4/11/16

Date _____



BRIGHTON ANALYTICAL, LLC

QUALITY ASSURANCE/QUALITY
CONTROL

ICP-MS

EPA METHOD 200.8/6020

REPRESENTATIVE BATCH PRECISION AND ACCURACY QUALITY CONTROL SUMMARY

Analysis Date: 4/8/2016	Standard ID: 033116 H2O	Batch: 4/6/2016 W3
Matrix Spike Lab ID: CC09095	Matrix: Total	Analyst: LT

Metals	Matrix Spike - Precision *			Matrix Spike - Accuracy**				Miscellaneous***		
	Matrix Spike (ug/L)	Matrix Spike Dup (ug/L)	RPD (%)	Spk Conc (ug/L)	MS Recovery (%)	MSD Recovery (%)	Sample Conc (ug/L)	Method Blk (ug/L)	LCS-Method STD (%)	Ind. Std. SPEX 1&3 (%)
Magnesium	16014	15781	1.5	10000	90.9	88.6	6922	<100	92.2	90.1
Calcium	34432	34111	0.9	10000	92.8	89.6	25151	<100	95.3	93.4
Copper	951	938	1.4	1000	93.7	92.4	14	<1	96.4	96.0
Lead	906	897	1.0	1000	90.4	89.5	2	<1	92.3	91.1

* Matrix spike precision range +/- 20% RPD

** Matrix spike accuracy range +/- 30% recovery

*** LCS accuracy range +/- 15% recovery / Ind std accuracy range +/- 10% recovery

Comments: _____